

EU DECLARATION OF CONFORMITY

Manufacturer Name/Address: Ansell Healthcare Europe NV/SA
Riverside Business Park,
Block J, Boulevard International 55,
1070 Brussels,
Belgium

SRN Number: BE-MF-000000691

Risk Class: Class I

Intended Purpose: A non-sterile medical device intended as an examination/treatment glove and a protective barrier when worn on the hands of healthcare providers, during patient examination, or for other sanitary purposes. The device is used mainly as a two-way barrier to protect both the patient and wearer against contaminants. This is a single-use device.

EMDN Code and Description: T01020204 - Nitrile Examination/Treatment Glove

Basic UDI-DI: 5414566 MTNAF700 3C

Product Name(s):

Product Name	Size	Product Code	Market Regions
Micro-Touch® Nitrile Accelerator Free	XS	700101	EMEA
Micro-Touch® Nitrile Accelerator Free	S	700102	EMEA
Micro-Touch® Nitrile Accelerator Free	M	700103	EMEA
Micro-Touch® Nitrile Accelerator Free	L	700104	EMEA
Micro-Touch® Nitrile Accelerator Free	XL	700105	EMEA

Conformity Assessment Procedure: Annex I & Annex II + Annex III

This EU Declaration of Conformity is issued under the sole responsibility of Ansell Healthcare Europe NV/SA.

We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

Signed on behalf of Ansell Healthcare Europe NV



Ansell Healthcare Europe NV
Riverside Business Park - Block J
Bld Internationalelaan 55
B-1070 Brussels
BELGIUM

Name: Samantha Marshall
Position: Director Regulatory Affairs Medical EMEA / APAC
Date of issue: 02 March 2023
Place of issue: Nuneaton, England
Version No: MED\MTNITACCL\003